

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL051921\
 Data File : VL037005.D
 Acq On : 19 May 2021 23:48
 Operator : SY/AP
 Sample : M2424-07DL 10X
 Misc : 400mL/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 KSV-06DL

Manual Integrations
 APPROVED

MMDadoda
 5/24/2021 3:02:13 PM

Quant Time: May 22 22:41:46 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL051821AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Tue May 18 2021
 QLast Update : Tue May 18 12:00:15 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.64	49	1130969	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.15	114	3264200	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.06	117	2995748	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.38	95	2363500	10.50	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	105.00%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
9) Propene	2.99	41	30022	0.405	ppbv	91
13) Ethanol	3.60	45	41263	4.168	ppbv	91
15) Acetone	3.83	43	952547	8.223	ppbv	98
20) Methylene Chloride	4.33	84	191174	3.079	ppbv #	85
26) Hexane	5.66	57	170811	1.201	ppbv #	44
29) Chloroform	5.73	83	53627m	0.267	ppbv	
32) 1,1,1-Trichloroethane	6.48	97	184219	0.854	ppbv	98
34) 2-Butanone	5.23	43	69635m	0.586	ppbv	
52) Tetrachloroethene	11.19	164	12069m	0.108	ppbv	
53) Toluene	9.77	91	34436	0.103	ppbv	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL051921\
Data File : VL037005.D
Acq On : 19 May 2021 23:48
Operator : SY/AP
Sample : M2424-07DL 10X
Misc : 400mL/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
KSV-06DL

Manual Integrations
APPROVED

MMDadoda
5/24/2021 3:02:13 PM

Quant Time: May 22 22:41:46 2021
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL051821AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LTue May
QLast Update : Tue May 18 12:00:15 2021
Response via : Initial Calibration

