

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL052124\
 Data File : VL041262.D
 Acq On : 21 May 2024 12:53
 Operator : SY/MD
 Sample : QC051524 CAN 10331
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 QC051524 CAN 10331

Quant Time: May 22 05:43:21 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL042924AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Wed May 22 05:39:49 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.198	49	918196	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	6.655	114	2476753	10.000	ppbv	0.00
55) Chlorobenzene-d5	11.462	117	2274490	10.000	ppbv	0.00

System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	13.774	95	1713031	9.976	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	99.800%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL052124\
Data File : VL041262.D
Acq On : 21 May 2024 12:53
Operator : SY/MD
Sample : QC051524 CAN 10331
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
QC051524 CAN 10331

Quant Time: May 22 05:43:21 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL042924AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Wed May 22 05:39:49 2024
Response via : Initial Calibration

