

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL052124\
 Data File : VL041275.D
 Acq On : 21 May 2024 21:38
 Operator : SY/MD
 Sample : CAN 10602
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 CAN 10602

Quant Time: May 22 05:48:46 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL042924AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Wed May 22 05:39:49 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.201	49	981527	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	6.658	114	2529969	10.000	ppbv	0.00
55) Chlorobenzene-d5	11.468	117	2329280	10.000	ppbv	0.00

System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	13.777	95	1804806	10.263	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	102.600%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL052124\
Data File : VL041275.D
Acq On : 21 May 2024 21:38
Operator : SY/MD
Sample : CAN 10602
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
CAN 10602

Quant Time: May 22 05:48:46 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL042924AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Wed May 22 05:39:49 2024
Response via : Initial Calibration

