

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL052225\
 Data File : VL042575.D
 Acq On : 22 May 2025 11:27
 Operator : SY/MD
 Sample : MDL 0.03
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 MDL 0.03

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 05/23/2025
 Supervised By :Mahesh Dadoda 05/23/2025

Quant Time: May 23 01:47:11 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL052025AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2
 QLast Update : Wed May 21 00:21:29 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.797	49	125849	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.975	114	388927	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.898	117	331913	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	10.390	95	257942	9.710	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	97.100%
Target Compounds						
5) Vinyl Chloride	1.586	62	341	0.041	ppbv	Qvalue 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL052225\
Data File : VL042575.D
Acq On : 22 May 2025 11:27
Operator : SY/MD
Sample : MDL 0.03
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
MDL 0.03

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 05/23/2025
Supervised By :Mahesh Dadoda 05/23/2025

Quant Time: May 23 01:47:11 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL052025AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Wed May 21 00:21:29 2025
Response via : Initial Calibration

