

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL052225\
 Data File : VL042576.D
 Acq On : 22 May 2025 11:59
 Operator : SY/MD
 Sample : MDL 0.03
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 MDL 0.03

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 05/23/2025
 Supervised By :Mahesh Dadoda 05/23/2025

Quant Time: May 23 01:48:06 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL052025AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2
 QLast Update : Wed May 21 00:21:29 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.800	49	119343	10.000	ppbv	0.01
33) 1,4-Difluorobenzene	3.978	114	355231	10.000	ppbv	0.01
55) Chlorobenzene-d5	8.901	117	304373	10.000	ppbv	0.01
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	10.393	95	239767	9.843	ppbv	0.01
Spiked Amount	10.000	Range	65 - 135	Recovery	=	98.400%
Target Compounds						
5) Vinyl Chloride	1.583	62	338m	0.043	ppbv	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL052225\
Data File : VL042576.D
Acq On : 22 May 2025 11:59
Operator : SY/MD
Sample : MDL 0.03
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1


Instrument :
MSVOA_L
ClientSampleId :
MDL 0.03

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 05/23/2025
Supervised By :Mahesh Dadoda 05/23/2025

Quant Time: May 23 01:48:06 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL052025AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Wed May 21 00:21:29 2025
Response via : Initial Calibration









#1
Bromochloromethane
Concen: 10.000 ppbv
RT: 2.800 min Scan# 81
Delta R.T. 0.010 min
Lab File: VL042576.D
Acq: 22 May 2025 11:59

Instrument :
MSVOA_L
ClientSampleId :
MDL 0.03

Tgt Ion: 49 Resp: 11934
Ion Ratio Lower Upper
49 100
128 55.4 28.6 85.8
130 71.0 36.6 109.8




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 05/23/2025
Supervised By :Mahesh Dadoda 05/23/2025



#5
Vinyl Chloride
Concen: 0.043 ppbv m
RT: 1.583 min Scan# 462
Delta R.T. 0.003 min
Lab File: VL042576.D
Acq: 22 May 2025 11:59

Tgt Ion: 62 Resp: 338
Ion Ratio Lower Upper
62 100
64 25.4 25.7 38.5#





#33

1,4-Difluorobenzene

Concen: 10.000 ppbv

RT: 3.978 min Scan# 1198

Delta R.T. 0.013 min

Lab File: VL042576.D

Acq: 22 May 2025 11:59

Instrument :

MSVOA_L

ClientSampleId :

MDL 0.03

Tgt Ion:114 Resp: 35523

Ion Ratio Lower Upper

114 100

63 24.0 18.4 27.6





88 17.5 14.0 21.0

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 05/23/2025

Supervised By :Mahesh Dadoda 05/23/2025



#55

Chlorobenzene-d5

Concen: 10.000 ppbv

RT: 8.901 min Scan# 2723

Delta R.T. 0.013 min

Lab File: VL042576.D

Acq: 22 May 2025 11:59




Tgt Ion:117 Resp: 304373


Ion Ratio Lower Upper

117 100

82 62.5 51.0 76.4


119 32.5 27.0 40.4





#68
 1-Bromo-4-Fluorobenzene
 Concen: 9.843 ppbv
 RT: 10.393 min Scan# 31
 Delta R.T. 0.013 min
 Lab File: VL042576.D
 Acq: 22 May 2025 11:59

Instrument :
 MSVOA_L
 ClientSampleId :
 MDL 0.03



Tgt Ion: 95 Resp: 23976
 Ion Ratio Lower Upper
 95 100
 174 65.6 52.0 78.0
 175 4.6 3.8 5.6

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 05/23/2025
 Supervised By :Mahesh Dadoda 05/23/2025

