

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL052319\
 Data File : VL033769.D
 Acq On : 23 May 2019 10:25
 Operator : SY/AP
 Sample : VSTDCCC010
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 LabSampleId :
 VSTDCCC010

Quant Time: May 24 04:42:10 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL051719AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri May 17 08:15:41 2019
 QLast Update : Fri May 17 08:15:41 2019
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Bromochloromethane	1.000	1.000	0.0	88	0.00
2 T	Dichlorodifluoromethane	2.107	1.669	20.8	81	0.00
3	Chlorodifluoromethane	1.145	1.180	-3.1	96	0.00
4	Chloromethane	0.655	0.673	-2.7	94	0.00
5 T	Vinyl Chloride	0.667	0.684	-2.5	97	0.00
6 T	Bromomethane	0.407	0.435	-6.9	95	0.00
7	Chloroethane	0.237	0.255	-7.6	98	0.00
8 T	Dichlorotetrafluoroethane	1.699	1.755	-3.3	97	0.00
9 T	Propene	0.487	0.509	-4.5	92	0.00
10 T	Heptane	1.380	1.440	-4.3	93	0.00
11 T	Trichlorofluoromethane	2.011	2.086	-3.7	92	0.00
12 T	1,1,2-Trichlorotrifluoroeth	1.405	1.408	-0.2	90	0.00
13	Ethanol	0.157	0.084	46.5#	60	0.00
14 T	Bromoethene	0.516	0.564	-9.3	99	0.00
15 T	Acetone	1.387	1.292	6.8	90	0.00
16 T	1,3-Butadiene	0.553	0.604	-9.2	99	0.00
17	tert-Butyl alcohol	1.122	1.089	2.9	82	0.00
18 T	1,1-Dichloroethene	0.593	0.620	-4.6	91	0.00
19 T	Isopropyl Alcohol	0.895	0.774	13.5	74	0.00
20 T	Methylene Chloride	0.534	0.529	0.9	94	0.00
21 T	Allyl Chloride	0.865	0.906	-4.7	89	0.00
22 T	trans-1,2-Dichloroethene	0.608	0.642	-5.6	97	0.00
23 T	Vinyl Acetate	1.829	1.986	-8.6	93	0.00
24 T	1,1-Dichloroethane	1.391	1.405	-1.0	95	0.00
25 T	Ethyl Acetate	2.437	2.429	0.3	91	0.00
26 T	Hexane	1.048	1.059	-1.0	89	0.00
27 T	Carbon Disulfide	1.438	1.586	-10.3	91	0.00
28 T	Methyl tert-Butyl Ether	1.708	1.769	-3.6	91	0.00
29 T	Chloroform	1.930	1.821	5.6	91	0.00
30 T	Cyclohexane	0.881	0.897	-1.8	87	0.00
31 T	cis-1,2-Dichloroethene	1.039	1.097	-5.6	96	0.00
32 T	1,1,1-Trichloroethane	1.927	1.857	3.6	87	0.00
33 I	1,4-Difluorobenzene	1.000	1.000	0.0	87	0.00
34 T	2-Butanone	0.620	0.667	-7.6	93	0.00
35 T	Carbon Tetrachloride	0.801	0.851	-6.2	90	0.00
36 T	Benzene	0.910	0.926	-1.8	86	0.00
37 T	1,2-Dichloroethane	0.587	0.600	-2.2	90	0.00
38 T	Trichloroethene	0.338	0.395	-16.9	94	0.00
39 T	1,2-Dichloropropane	0.339	0.357	-5.3	90	0.00
40 T	1,4-Dioxane	0.142	0.131	7.7	78	0.00
41 T	Tetrahydrofuran	0.347	0.370	-6.6	86	0.00
42 T	Bromodichloromethane	0.803	0.889	-10.7	95	0.00
43	Methyl Methacrylate	0.335	0.395	-17.9	94	0.00
44 T	2,2,4-Trimethylpentane	1.537	1.666	-8.4	93	0.00
45 T	t-1,3-Dichloropropene	0.424	0.506	-19.3	86	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL052319\
 Data File : VL033769.D
 Acq On : 23 May 2019 10:25
 Operator : SY/AP
 Sample : VSTDCCC010
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 LabSampleId :
 VSTDCCC010

Quant Time: May 24 04:42:10 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL051719AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri May 17 08:15:41 2019
 QLast Update : Fri May 17 08:15:41 2019
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	cis-1,3-Dichloropropene	0.497	0.571	-14.9	85	0.00
47 T	1,1,2-Trichloroethane	0.385	0.395	-2.6	89	0.00
48 T	Dibromochloromethane	0.668	0.721	-7.9	86	0.00
49 T	Bromoform	0.562	0.692	-23.1	99	0.00
50 T	4-Methyl-2-Pentanone	0.953	1.010	-6.0	86	0.00
51 T	2-Hexanone	0.744	0.831	-11.7	86	0.00
52 T	Tetrachloroethene	0.320	0.355	-10.9	86	0.00
53 T	Toluene	0.981	1.095	-11.6	87	0.00
54 T	1,2-Dibromoethane	0.588	0.617	-4.9	87	0.00
55 I	Chlorobenzene-d5	1.000	1.000	0.0	95	0.00
56	1,1,1,2-Tetrachloroethane	0.463	0.457	1.3	94	-0.01
57 T	Chlorobenzene	0.823	0.800	2.8	94	0.00
58 T	Ethyl Benzene	1.324	1.430	-8.0	93	0.00
59 T	m/p-Xylene	1.158	1.253	-8.2	98	0.00
60 T	o-Xylene	1.167	1.227	-5.1	96	0.00
61 T	Styrene	0.786	0.915	-16.4	95	0.00
62	Isopropylbenzene	1.572	1.624	-3.3	95	0.00
63 T	1,1,2,2-Tetrachloroethane	0.849	0.868	-2.2	99	0.00
64	n-propylbenzene	0.407	0.421	-3.4	94	0.00
65	tert-Butylbenzene	1.435	1.498	-4.4	94	0.00
66 T	Benzyl Chloride	0.683	0.765	-12.0	86	0.00
67	sec-Butylbenzene	2.008	2.083	-3.7	94	0.00
68 S	1-Bromo-4-Fluorobenzene	0.811	0.786	3.1	92	0.00
69	p-Isopropyltoluene	1.660	1.708	-2.9	91	0.00
70	n-Butylbenzene	1.780	1.760	1.1	90	0.00
71	2-Chlorotoluene	1.166	1.234	-5.8	94	0.00
72 T	4-Ethyltoluene	1.334	1.424	-6.7	93	0.00
73 T	1,3,5-Trimethylbenzene	1.302	1.363	-4.7	95	0.00
74 T	1,2,4-Trimethylbenzene	1.417	1.431	-1.0	93	0.00
75 T	1,3-Dichlorobenzene	0.868	0.824	5.1	92	0.00
76 T	1,4-Dichlorobenzene	0.825	0.807	2.2	91	0.00
77 T	1,2-Dichlorobenzene	0.816	0.793	2.8	92	0.00
78 T	Hexachloro-1,3-Butadiene	0.617	0.590	4.4	100	0.00
79 T	Naphthalene	1.316	1.194	9.3	76	0.00
80 T	Naphthalene,2-methyl-	0.500	0.531	-6.2	75	0.00
81 T	1,2,4-Trichlorobenzene	0.771	0.673	12.7	79	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0