

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL052919\
 Data File : VL033853.D
 Acq On : 30 May 2019 2:53
 Operator : SY/AP
 Sample : K3067-05DL 10X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 SG-7-20190523-ODL

Manual Integrations
 APPROVED

apatel
 5/31/2019 12:45:51 PM

Quant Time: May 30 06:28:26 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL051719AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri May 1
 QLast Update : Fri May 17 08:15:41 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.75	49	638499	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.28	114	1453933	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.23	117	1565765	10.00	ppbv	0.00

System Monitoring Compounds		R.T.	QIon	Response	Conc	Units	Dev(Min)
68) 1-Bromo-4-Fluorobenzene	14.57	95	1259002	9.92	ppbv	0.00	
Spiked Amount	10.000	Range	65 - 135	Recovery	=	99.20%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
13) Ethanol	3.66	45	17718m	1.762	ppbv	
15) Acetone	3.91	43	76781	0.867	ppbv	100
20) Methylene Chloride	4.41	84	8532	0.250	ppbv #	85
26) Hexane	5.77	57	8928	0.133	ppbv #	42
29) Chloroform	5.84	83	13879	0.113	ppbv	96
34) 2-Butanone	5.34	43	10609m	0.118	ppbv	
52) Tetrachloroethene	11.37	164	30109	0.648	ppbv	100
53) Toluene	9.94	91	18614m	0.130	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL052919\
Data File : VL033853.D
Acq On : 30 May 2019 2:53
Operator : SY/AP
Sample : K3067-05DL 10X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
SG-7-20190523-ODL

Manual Integrations
APPROVED

apatel
5/31/2019 12:45:51 PM

Quant Time: May 30 06:28:26 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL051719AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri May
QLast Update : Fri May 17 08:15:41 2019
Response via : Initial Calibration

