

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL052919\
 Data File : VL033870.D
 Acq On : 30 May 2019 14:35
 Operator : SY/AP
 Sample : K3067-09DL DUP 8X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 SG-23-20190523-0

Manual Integrations
 APPROVED

apatel
 5/31/2019 2:51:02 PM

Quant Time: May 31 08:14:44 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL051719AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri May 1
 QLast Update : Fri May 17 08:15:41 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.75	49	651861	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.28	114	1450120	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.23	117	1461296	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.57	95	1196466	10.10	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	101.00%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
13) Ethanol	3.67	45	22379m	2.180	ppbv	
15) Acetone	3.91	43	74787	0.827	ppbv	96
19) Isopropyl Alcohol	4.05	45	22921m	0.393	ppbv	
26) Hexane	5.77	57	80685	1.181	ppbv #	38
32) 1,1,1-Trichloroethane	6.61	97	13085m	0.104	ppbv	
52) Tetrachloroethene	11.36	164	63404	1.367	ppbv	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL052919\
Data File : VL033870.D
Acq On : 30 May 2019 14:35
Operator : SY/AP
Sample : K3067-09DLUP 8X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
SG-23-20190523-0

Manual Integrations
APPROVED

apatel
5/31/2019 2:51:02 PM

Quant Time: May 31 08:14:44 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL051719AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri May
QLast Update : Fri May 17 08:15:41 2019
Response via : Initial Calibration

