

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL052925\
 Data File : VL042593.D
 Acq On : 29 May 2025 19:32
 Operator : SY/MD
 Sample : Q2124-02DL 20X
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 SV2DL

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 05/30/2025
 Supervised By :Mahesh Dadoda 05/30/2025

Quant Time: May 30 02:13:13 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL052925AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2025
 QLast Update : Fri May 30 02:01:56 2025
 Response via : Initial Calibration

Compound		R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards							
1) Bromochloromethane		2.797	49	107351	10.000	ppbv	0.00
33) 1,4-Difluorobenzene		3.972	114	309485	10.000	ppbv	0.00
55) Chlorobenzene-d5		8.895	117	273119	10.000	ppbv	0.00
System Monitoring Compounds							
68) 1-Bromo-4-Fluorobenzene		10.387	95	204804	10.003	ppbv	0.00
Spiked Amount		10.000	Range 65 - 135		Recovery	= 100.000%	
Target Compounds							
							Qvalue
13) Ethanol		1.729	45	581m	0.666	ppbv	
15) Acetone		1.842	43	4823	0.373	ppbv	95
20) Methylene Chloride		2.065	84	1149	0.269	ppbv #	87
26) Hexane		2.836	57	7026	0.437	ppbv #	42
27) Carbon Disulfide		2.156	76	7589	0.615	ppbv #	1
29) Chloroform		2.862	83	6940	0.332	ppbv #	82
34) 2-Butanone		2.570	43	4804	0.218	ppbv	100
38) Trichloroethene		4.574	130	389m	0.030	ppbv	
52) Tetrachloroethene		8.241	164	2983m	0.262	ppbv	
53) Toluene		6.953	91	4812m	0.137	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL052925\
Data File : VL042593.D
Acq On : 29 May 2025 19:32
Operator : SY/MD
Sample : Q2124-02DL 20X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
SV2DL

Quant Time: May 30 02:13:13 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL052925AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Fri May 30 02:01:56 2025
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 05/30/2025
Supervised By :Mahesh Dadoda 05/30/2025

