

Data Path : Z:\voasrv\HPCHEM1\MSVOA L\Data\VL053118\
 Data File : VL032021.D
 Acq On : 31 May 2018 19:23
 Operator : SY/AP
 Sample : VSTDICC0.03
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC0.03

Manual Integrations
 APPROVED

MMDadoda
 6/4/2018 10:54:58 AM

Quant Time: Jun 01 02:28:13 2018
 Quant Method : W:\HPCHEM1\MSVOA L\METHODS\VL053118AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Jun 01 02:28:13 2018
 QLast Update : Fri Jun 01 01:58:27 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.86	49	1389112	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.40	114	2428370	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.36	117	2438527	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.70	95	1903914	10.05	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	100.50%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
5) Vinyl Chloride	3.37	62	2244m	0.03	ppbv	
32) 1,1,1-Trichloroethane	6.73	97	7419m	0.03	ppbv	
35) Carbon Tetrachloride	7.25	117	7674m	0.04	ppbv	
38) Trichloroethene	8.08	130	2306m	0.03	ppbv	
52) Tetrachloroethene	11.50	164	2843m	0.03	ppbv	
63) 1,1,2,2-Tetrachloroethane	13.96	83	8615m	0.04	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA L\Data\VL053118\
Data File : VL032021.D
Acq On : 31 May 2018 19:23
Operator : SY/AP
Sample : VSTDICC0.03
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VSTDICC0.03

Manual Integrations
APPROVED

MMDadoda
6/4/2018 10:54:58 AM

Quant Time: Jun 01 02:28:13 2018
Quant Method : W:\HPCHEM1\MSVOA L\METHODS\VL053118AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Jun
QLast Update : Fri Jun 01 01:58:27 2018
Response via : Initial Calibration

