

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL060121\
 Data File : VL037078.D
 Acq On : 1 Jun 2021 20:14
 Operator : SY/AP
 Sample : M2559-04DL2 40X
 Misc : 400mL/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 HSG4DL2

Manual Integrations
 APPROVED

MMDadoda
 6/7/2021 2:59:08 PM

Quant Time: Jun 02 04:04:34 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL051821AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Tue May 18 2021
 QLast Update : Tue May 18 12:00:15 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.65	49	1498693	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.16	114	4569890	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.06	117	4121849	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.39	95	2961213	9.56	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	95.60%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
9) Propene	3.01	41	13043m	0.133	ppbv	
13) Ethanol	3.63	45	40032m	3.052	ppbv	
15) Acetone	3.85	43	313632	2.043	ppbv	91
20) Methylene Chloride	4.33	84	165811	2.015	ppbv	96
26) Hexane	5.67	57	131883	0.700	ppbv #	46

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL060121\
Data File : VL037078.D
Acq On : 1 Jun 2021 20:14
Operator : SY/AP
Sample : M2559-04DL2 40X
Misc : 400mL/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
HSG4DL2

Manual Integrations
APPROVED

MMDadoda
6/7/2021 2:59:08 PM

Quant Time: Jun 02 04:04:34 2021
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL051821AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LTue May
QLast Update : Tue May 18 12:00:15 2021
Response via : Initial Calibration

