

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL060121\
 Data File : VL037089.D
 Acq On : 2 Jun 2021 3:33
 Operator : SY/AP
 Sample : M2548-05DL 10X
 Misc : 400mL/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 IA-2ADL

Manual Integrations
 APPROVED

MMDadoda
 6/7/2021 2:59:30 PM

Quant Time: Jun 06 21:35:13 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL051821AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Tue May 18 2021
 QLast Update : Tue May 18 12:00:15 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.65	49	1473412	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.16	114	4295478	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.06	117	3881023	10.00	ppbv	0.00

System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	14.40	95	2736941	9.39	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	93.90%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
2) Dichlorodifluoromethane	3.04	85	30958	0.124	ppbv	97
13) Ethanol	3.59	45	793508m	61.529	ppbv	
15) Acetone	3.84	43	486665	3.225	ppbv	91
19) Isopropyl Alcohol	3.97	45	85529m	1.054	ppbv	
20) Methylene Chloride	4.33	84	36173	0.447	ppbv	89
53) Toluene	9.79	91	46218m	0.105	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL060121\
Data File : VL037089.D
Acq On : 2 Jun 2021 3:33
Operator : SY/AP
Sample : M2548-05DL 10X
Misc : 400mL/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
IA-2ADL

Manual Integrations
APPROVED

MMDadoda
6/7/2021 2:59:30 PM

Quant Time: Jun 06 21:35:13 2021
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL051821AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Tue May
QLast Update : Tue May 18 12:00:15 2021
Response via : Initial Calibration

