

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL060325\
 Data File : VL042604.D
 Acq On : 03 Jun 2025 10:53
 Operator : SY/MD
 Sample : VL0603ABL01
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VL0603ABL01

Quant Time: Jun 04 01:37:18 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL052925AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Fri May 30 02:01:56 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.784	49	99329	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.953	114	280531	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.879	117	249364	10.000	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	10.371	95	187874	10.051	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	100.500%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL060325\
Data File : VL042604.D
Acq On : 03 Jun 2025 10:53
Operator : SY/MD
Sample : VL0603ABL01
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VL0603ABL01

Quant Time: Jun 04 01:37:18 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL052925AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Fri May 30 02:01:56 2025
Response via : Initial Calibration

