

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL060325\
 Data File : VL042611.D
 Acq On : 03 Jun 2025 15:12
 Operator : SY/MD
 Sample : Q2126-13 LOD 0.03
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 LOD-MDL-AIR-01-QT2-2025

Quant Time: Jun 04 01:52:34 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL052925AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2
 QLast Update : Fri May 30 02:01:56 2025
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 06/04/2025
 Supervised By :Mahesh Dadoda 06/06/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.787	49	97231	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.958	114	264710	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.881	117	232229	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	10.377	95	177712	10.208	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	102.100%
Target Compounds						
						Qvalue
32) 1,1,1-Trichloroethane	3.366	97	777m	0.040	ppbv	
52) Tetrachloroethene	8.228	164	402m	0.041	ppbv	
63) 1,1,2,2-Tetrachloroethane	9.966	83	805m	0.041	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL060325\
Data File : VL042611.D
Acq On : 03 Jun 2025 15:12
Operator : SY/MD
Sample : Q2126-13 LOD 0.03
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
LOD-MDL-AIR-01-QT2-2025

Quant Time: Jun 04 01:52:34 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL052925AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Fri May 30 02:01:56 2025
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : Semsettin Yesilyurt 06/04/2025
Supervised By : Mahesh Dadoda 06/06/2025

