

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL061319\
 Data File : VL033918.D
 Acq On : 14 Jun 2019 4:48
 Operator : SY/AP
 Sample : K3367-02DL 10X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 IA2-2NDFLRDL

Quant Time: Jun 14 09:14:44 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL060719AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Jun 07 08:53:27 2019
 QLast Update : Fri Jun 07 08:53:27 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.75	49	890569	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.27	114	2020801	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.22	117	1882132	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.56	95	1564818	10.12	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	101.20%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
2) Dichlorodifluoromethane	3.07	85	14205	0.098	ppbv	98
9) Propene	3.04	41	14910	0.380	ppbv	88
13) Ethanol	3.64	45	1764303	138.230	ppbv	97
15) Acetone	3.89	43	218330	1.681	ppbv	97
19) Isopropyl Alcohol	4.02	45	35217	0.475	ppbv #	86
52) Tetrachloroethene	11.35	164	51849	0.955	ppbv	90
53) Toluene	9.92	91	23783	0.139	ppbv	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL061319\
Data File : VL033918.D
Acq On : 14 Jun 2019 4:48
Operator : SY/AP
Sample : K3367-02DL 10X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
IA2-2NDFLRDL

Quant Time: Jun 14 09:14:44 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL060719AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Jun 07 08:53:27 2019
QLast Update : Fri Jun 07 08:53:27 2019
Response via : Initial Calibration

