

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL061319\
 Data File : VL033920.D
 Acq On : 14 Jun 2019 6:07
 Operator : SY/AP
 Sample : K3367-03DL 10X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 OA1DL

Manual Integrations
 APPROVED

MMDadoda
 6/17/2019 12:37:35 AM

Quant Time: Jun 14 09:32:28 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL060719AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Jun 0
 QLast Update : Fri Jun 07 08:53:27 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.75	49	792971	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.27	114	1775285	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.22	117	1916550	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.56	95	1402894	8.91	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	89.10%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
13) Ethanol	3.65	45	38396m	3.379	ppbv	
15) Acetone	3.90	43	53161	0.460	ppbv	100
20) Methylene Chloride	4.40	84	21034	0.466	ppbv #	84
26) Hexane	5.77	57	13644m	0.170	ppbv	
53) Toluene	9.93	91	42350m	0.282	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL061319\
Data File : VL033920.D
Acq On : 14 Jun 2019 6:07
Operator : SY/AP
Sample : K3367-03DL 10X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
QA1DL

Manual Integrations
APPROVED

MMDadoda
6/17/2019 12:37:35 AM

Quant Time: Jun 14 09:32:28 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL060719AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Jun
QLast Update : Fri Jun 07 08:53:27 2019
Response via : Initial Calibration

