

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL061319\
Data File : VL033903.D
Acq On : 13 Jun 2019 18:43
Operator : SY/AP
Sample : OC061219 CAN 10490
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
QC061219 CAN 10490

Quant Time: Jun 14 06:57:24 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL060719AIR.M
Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Fri Jun 07 08:53:27 2019
QLast Update : Fri Jun 07 08:53:27 2019
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.75	49	963280	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.27	114	2198862	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.22	117	2021067	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.56	95	1672327	10.07	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	100.70%


Target Compounds	Ovalue
------------------	--------


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL061319\
Data File : VL033903.D
Acq On : 13 Jun 2019 18:43
Operator : SY/AP
Sample : OC061219 CAN 10490
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
QC061219 CAN 10490

Quant Time: Jun 14 06:57:24 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL060719AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Jun 07 08:53:27 2019
QLast Update : Fri Jun 07 08:53:27 2019
Response via : Initial Calibration





#1

Bromochloromethane

Concen: 10.000 ppbv

RT: 5.75 min Scan# 874

Delta R.T. -0.01 min

Lab File: VL033903.D

Acq: 13 Jun 2019 18:43

Instrument :

MSVOA_L

ClientSampleId :

QC061219 CAN 10490


Tot Ion: 49 Resp: 963280

Ion Ratio Lower Upper

49 100

128 51.7 25.5 76.5




130 66.2 32.5 97.4



Abundance Ion 49.10 (48.80 to 49.80): VL0

Ion 128.00 (127.70 to 128.70): V

Ion 130.00 (129.70 to 130.70): V



#33

1,4-Difluorobenzene

Concen: 10.000 ppbv

RT: 7.27 min Scan# 1349

Delta R.T. -0.01 min

Lab File: VL033903.D

Acq: 13 Jun 2019 18:43


Tgt Ion:114 Resp: 2198862

Ion Ratio Lower Upper

114 100

63 26.5 21.1 31.7



88 18.1 14.2 21.4




Abundance Ion 114.10 (113.80 to 114.80): V

Ion 63.10 (62.80 to 63.80): VL0

Ion 88.10 (87.80 to 88.80): VL0









#55
Chlorobenzene-d5
Concen: 10.000 ppbv
RT: 12.22 min Scan# 2888
Delta R.T. -0.00 min
Lab File: VL033903.D
Acq: 13 Jun 2019 18:43

Instrument :
MSVOA_L
ClientSampleId :
QC061219 CAN 10490

Tot Ion:117 Resp: 2021067
Ion Ratio Lower Upper
117 100
82 69.0 50.2 75.2
119 32.7 24.8 37.2



#68
1-Bromo-4-Fluorobenzene
Concen: 10.074 ppbv
RT: 14.56 min Scan# 3614
Delta R.T. -0.01 min
Lab File: VL033903.D
Acq: 13 Jun 2019 18:43

Tgt Ion: 95 Resp: 1672327
Ion Ratio Lower Upper
95 100
174 66.2 54.0 81.0
175 4.7 3.8 5.8

