

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL061621\
Data File : VL037172.D
Acq On : 16 Jun 2021 16:01
Operator : SY/AP
Sample : VL0616ABL01
Misc : 400mL/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VL0616ABL01

Quant Time: Jun 26 00:11:51 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL061621AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LWed Jun 16 14
QLast Update : Wed Jun 16 14:06:32 2021
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.67	49	1451813	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.18	114	4310115	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.10	117	3761208	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.42	95	2869578	10.07	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	100.70%

Target Compounds	Ovalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL061621\
Data File : VL037172.D
Acq On : 16 Jun 2021 16:01
Operator : SY/AP
Sample : VL0616ABL01
Misc : 400mL/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VL0616ABL01

Quant Time: Jun 26 00:11:51 2021
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL061621AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Jun 16 1
QLast Update : Wed Jun 16 14:06:32 2021
Response via : Initial Calibration

