

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL061625\
 Data File : VL042635.D
 Acq On : 16 Jun 2025 18:38
 Operator : SY/MD
 Sample : QC041825 CAN 33628
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 QC041825 CAN 33628

Quant Time: Jun 17 03:09:22 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL052925AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Fri May 30 02:01:56 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.790	49	95799	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.965	114	251218	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.888	117	221734	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	10.383	95	167623	10.085	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	100.800%
Target Compounds						
					Qvalue	
15) Acetone	1.839	43	2708	0.235	ppbv #	87
20) Methylene Chloride	2.065	84	2039	0.535	ppbv	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL061625\
Data File : VL042635.D
Acq On : 16 Jun 2025 18:38
Operator : SY/MD
Sample : QC041825 CAN 33628
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
QC041825 CAN 33628

Quant Time: Jun 17 03:09:22 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL052925AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Fri May 30 02:01:56 2025
Response via : Initial Calibration

