

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL061625\
 Data File : VL042638.D
 Acq On : 16 Jun 2025 20:21
 Operator : SY/MD
 Sample : QC061625 CAN 23182
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 QC061625 CAN 23182

Quant Time: Jun 17 03:10:35 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL052925AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Fri May 30 02:01:56 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.790	49	97935	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.962	114	250477	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.888	117	219985	10.000	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	10.380	95	172841	10.481	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	104.800%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL061625\
Data File : VL042638.D
Acq On : 16 Jun 2025 20:21
Operator : SY/MD
Sample : QC061625 CAN 23182
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
QC061625 CAN 23182

Quant Time: Jun 17 03:10:35 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL052925AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Fri May 30 02:01:56 2025
Response via : Initial Calibration









#1
Bromochloromethane
Concen: 10.000 ppbv
RT: 2.790 min Scan# 835
Delta R.T. 0.000 min
Lab File: VL042638.D
Acq: 16 Jun 2025 20:21




Instrument :
MSVOA_L
ClientSampleId :
QC061625 CAN 23182


Tgt Ion: 49 Resp: 97935
Ion Ratio Lower Upper
49 100
128 46.3 24.2 72.6
130 59.0 30.4 91.2



#33
1,4-Difluorobenzene
Concen: 10.000 ppbv
RT: 3.962 min Scan# 1197
Delta R.T. 0.000 min
Lab File: VL042638.D
Acq: 16 Jun 2025 20:21

Tgt Ion: 114 Resp: 250477
Ion Ratio Lower Upper
114 100
63 26.6 19.0 28.6
88 18.1 13.6 20.4









#55
Chlorobenzene-d5
Concen: 10.000 ppbv
RT: 8.888 min Scan# 21
Delta R.T. 0.000 min
Lab File: VL042638.D
Acq: 16 Jun 2025 20:21

Instrument :
MSVOA_L
ClientSampleId :
QC061625 CAN 23182

Tgt Ion:117 Resp: 219985
Ion Ratio Lower Upper
117 100
82 66.6 51.2 76.8
119 32.8 26.3 39.5



#68
1-Bromo-4-Fluorobenzene
Concen: 10.481 ppbv
RT: 10.380 min Scan# 3180
Delta R.T. 0.000 min
Lab File: VL042638.D
Acq: 16 Jun 2025 20:21

Tgt Ion: 95 Resp: 172841
Ion Ratio Lower Upper
95 100
174 61.5 53.1 79.7
175 4.5 4.2 6.2

