

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL061925\
 Data File : VL042659.D
 Acq On : 19 Jun 2025 17:26
 Operator : SY/MD
 Sample : VIBLK
 Misc : 400mL/MSVOA_L
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VIBLK

Quant Time: Jun 20 01:40:59 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL052925AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Fri May 30 02:01:56 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.800	49	107257	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.978	114	276044	10.000	ppbv	0.02
55) Chlorobenzene-d5	8.901	117	239558	10.000	ppbv	0.01

System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	10.393	95	182169	10.144	ppbv	0.01
Spiked Amount	10.000	Range	65 - 135	Recovery	=	101.400%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL061925\
Data File : VL042659.D
Acq On : 19 Jun 2025 17:26
Operator : SY/MD
Sample : VIBLK
Misc : 400mL/MSVOA_L
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VIBLK

Quant Time: Jun 20 01:40:59 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL052925AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Fri May 30 02:01:56 2025
Response via : Initial Calibration

