

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL062722\
 Data File : VL039376.D
 Acq On : 28 Jun 2022 2:55
 Operator : SY/AP
 Sample : N3453-02DL 20X
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 SG-2DL

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 06/29/2022
 Supervised By :Mahesh Dadoda 06/30/2022

Quant Time: Jun 28 04:41:42 2022
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL053122AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LWed Jun 0
 QLast Update : Wed Jun 01 01:22:18 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.661	49	1076164	10.000	ppbv	-0.01
33) 1,4-Difluorobenzene	7.169	114	3047116	10.000	ppbv	-0.01
55) Chlorobenzene-d5	12.076	117	2643673	10.000	ppbv	-0.02
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	14.404	95	1946083	8.904	ppbv	-0.01
Spiked Amount	10.000	Range	65 - 135	Recovery	=	89.000%
Target Compounds						
						Qvalue
13) Ethanol	3.616	45	18541m	5.694	ppbv	
15) Acetone	3.867	43	96432m	0.860	ppbv	
53) Toluene	9.799	91	37192m	0.131	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL062722\
Data File : VL039376.D
Acq On : 28 Jun 2022 2:55
Operator : SY/AP
Sample : N3453-02DL 20X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
SG-2DL

Quant Time: Jun 28 04:41:42 2022
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL053122AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LWed Jun
QLast Update : Wed Jun 01 01:22:18 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 06/29/2022
Supervised By :Mahesh Dadoda 06/30/2022

