

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL062918\
 Data File : VL032150.D
 Acq On : 29 Jun 2018 21:14
 Operator : SY/AP
 Sample : VSTDICC0.03
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC0.03

Manual Integrations
 APPROVED

MMDadoda
 7/2/2018 4:25:07 PM

Quant Time: Jun 29 21:50:31 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL062918AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Jun 29 2018
 QLast Update : Fri Jun 29 20:36:18 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.85	49	1819357	10.00	ppbv	-0.02
33) 1,4-Difluorobenzene	7.38	114	4379837	10.00	ppbv	-0.02
55) Chlorobenzene-d5	12.34	117	4064155	10.00	ppbv	-0.02

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.68	95	3030498	10.03	ppbv	-0.02
Spiked Amount	10.000	Range	65 - 135	Recovery	=	100.30%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
5) Vinyl Chloride	3.36	62	3929	0.03	ppbv	# 76
32) 1,1,1-Trichloroethane	6.71	97	10917m	0.03	ppbv	
35) Carbon Tetrachloride	7.23	117	10616m	0.03	ppbv	
38) Trichloroethene	8.06	130	6070m	0.03	ppbv	
52) Tetrachloroethene	11.48	164	6098m	0.03	ppbv	
63) 1,1,2,2-Tetrachloroethane	13.94	83	17210m	0.04	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL062918\
Data File : VL032150.D
Acq On : 29 Jun 2018 21:14
Operator : SY/AP
Sample : VSTDICC0.03
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VSTDICC0.03

Manual Integrations
APPROVED

MMDadoda
7/2/2018 4:25:07 PM

Quant Time: Jun 29 21:50:31 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL062918AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Jun
QLast Update : Fri Jun 29 20:36:18 2018
Response via : Initial Calibration

