

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL062922\
 Data File : VL039414.D
 Acq On : 30 Jun 2022 3:08
 Operator : SY/AP
 Sample : N3530-01DL 10X
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 SV1DL

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 07/01/2022
 Supervised By :Mahesh Dadoda 07/01/2022

Quant Time: Jun 30 05:47:13 2022

Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL062922AIR.M

Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_LThu Jun 30 05:47:13 2022

QLast Update : Thu Jun 30 05:40:11 2022

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.655	49	1343328	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	7.160	114	3740788	10.000	ppbv	0.00
55) Chlorobenzene-d5	12.066	117	3243393	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	14.394	95	2432739	10.111	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	101.100%
Target Compounds						
					Qvalue	
9) Propene	3.018	41	28719m	0.373	ppbv	
13) Ethanol	3.610	45	14647m	3.280	ppbv	
15) Acetone	3.841	43	1785455	11.235	ppbv	97
17) tert-Butyl alcohol	4.298	59	46218	0.343	ppbv #	91
20) Methylene Chloride	4.340	84	26542	0.419	ppbv	97
26) Hexane	5.674	57	40857	0.311	ppbv #	34
34) 2-Butanone	5.240	43	84326	0.448	ppbv #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL062922\
Data File : VL039414.D
Acq On : 30 Jun 2022 3:08
Operator : SY/AP
Sample : N3530-01DL 10X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
SV1DL

Manual Integrations
APPROVED

Reviewed By : Semsettin Yesilyurt 07/01/2022
Supervised By : Mahesh Dadoda 07/01/2022

Quant Time: Jun 30 05:47:13 2022
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL062922AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu Jun
QLast Update : Thu Jun 30 05:40:11 2022
Response via : Initial Calibration

