

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL063020\
 Data File : VL035134.D
 Acq On : 30 Jun 2020 20:53
 Operator : SY/AP
 Sample : L3124-01DL 40X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 INFLUENTDL

Manual Integrations
 APPROVED

MMDadoda
 7/2/2020 5:00:47 PM

Quant Time: Jul 01 04:50:17 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL061920AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Sat Jun 20 2020
 QLast Update : Sat Jun 20 02:10:36 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.66	49	1252968	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.18	114	2939862	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.12	117	2752409	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.46	95	2321907	10.16	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	101.60%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
9) Propene	2.97	41	25679	0.469	ppbv	92
13) Ethanol	3.58	45	51700	2.497	ppbv	95
15) Acetone	3.83	43	169212	1.117	ppbv #	76
19) Isopropyl Alcohol	3.95	45	156499	1.591	ppbv #	97
20) Methylene Chloride	4.32	84	35058	0.572	ppbv	97
24) 1,1-Dichloroethane	5.00	63	28725m	0.185	ppbv	
26) Hexane	5.68	57	16035m	0.130	ppbv	
32) 1,1,1-Trichloroethane	6.51	97	158409	0.885	ppbv	99
34) 2-Butanone	5.24	43	412838	2.445	ppbv	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL063020\
Data File : VL035134.D
Acq On : 30 Jun 2020 20:53
Operator : SY/AP
Sample : L3124-01DL 40X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
INFLUENTDL

Manual Integrations
APPROVED

MMDadoda
7/2/2020 5:00:47 PM

Quant Time: Jul 01 04:50:17 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL061920AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Sat Jun
QLast Update : Sat Jun 20 02:10:36 2020
Response via : Initial Calibration

