

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL063020\
 Data File : VL035135.D
 Acq On : 30 Jun 2020 21:31
 Operator : SY/AP
 Sample : L3124-03DL 40X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 EFFLUENTDL

Manual Integrations
 APPROVED

MMDadoda
 7/2/2020 5:00:49 PM

Quant Time: Jul 01 14:47:32 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL061920AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Sat Jun 20 2020
 QLast Update : Sat Jun 20 02:10:36 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.66	49	1238454	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.18	114	2882728	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.12	117	2708768	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.45	95	2293507	10.19	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	101.90%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
9) Propene	2.97	41	83216	1.536	ppbv	100
13) Ethanol	3.58	45	36661m	1.791	ppbv	
15) Acetone	3.82	43	420618	2.809	ppbv #	59
24) 1,1-Dichloroethane	5.00	63	41404	0.270	ppbv	98
34) 2-Butanone	5.23	43	1505436	9.094	ppbv	100
51) 2-Hexanone	10.18	43	123536m	0.671	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL063020\
Data File : VL035135.D
Acq On : 30 Jun 2020 21:31
Operator : SY/AP
Sample : L3124-03DL 40X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
EFFLUENTDL

Manual Integrations
APPROVED

MMDadoda
7/2/2020 5:00:49 PM

Quant Time: Jul 01 14:47:32 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL061920AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Sat Jun
QLast Update : Sat Jun 20 02:10:36 2020
Response via : Initial Calibration

