

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL063020\
 Data File : VL035138.D
 Acq On : 30 Jun 2020 23:33
 Operator : SY/AP
 Sample : QC062420 CAN 10197
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 QC062420 CAN 10197

Manual Integrations
 APPROVED

MMDadoda
 7/2/2020 5:00:51 PM

Quant Time: Jul 01 05:05:07 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL061920AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Sat Jun 20 2020
 QLast Update : Sat Jun 20 02:10:36 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.66	49	1083711	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.18	114	2559128	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.12	117	2405169	10.00	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	14.46	95	2070625	10.36	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	103.60%
Target Compounds						
15) Acetone	3.84	43	32597m	0.249	ppbv	Ovalue
20) Methylene Chloride	4.33	84	74533	1.406	ppbv	# 82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL063020\
Data File : VL035138.D
Acq On : 30 Jun 2020 23:33
Operator : SY/AP
Sample : QC062420 CAN 10197
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
QC062420 CAN 10197

Manual Integrations
APPROVED

MMDadoda
7/2/2020 5:00:51 PM

Quant Time: Jul 01 05:05:07 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL061920AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Sat Jun
QLast Update : Sat Jun 20 02:10:36 2020
Response via : Initial Calibration

