

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL063020\
 Data File : VL035140.D
 Acq On : 1 Jul 2020 00:55
 Operator : SY/AP
 Sample : QC062920 CAN 10661
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 QC062920 CAN 10661

Manual Integrations
 APPROVED

MMDadoda
 7/2/2020 5:01:00 PM

Quant Time: Jul 01 05:09:56 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL061920AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Sat Jun 20 2020
 QLast Update : Sat Jun 20 02:10:36 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.66	49	1121723	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.18	114	2773238	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.12	117	2709821	10.00	ppbv	0.00

System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	14.46	95	2241106	9.96	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	99.60%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
5) Vinyl Chloride	3.22	62	11792	0.179	ppbv #	74
10) Heptane	8.14	43	85777	0.675	ppbv #	26
15) Acetone	3.84	43	329555	2.430	ppbv #	62
26) Hexane	5.68	57	537743	4.873	ppbv #	40
30) Cyclohexane	7.14	84	77287m	0.976	ppbv	
36) Benzene	6.90	78	127243	0.593	ppbv #	93
38) Trichloroethene	7.85	130	90622	1.086	ppbv	98
44) 2,2,4-Trimethylpentane	7.88	57	163135	0.435	ppbv #	57
52) Tetrachloroethene	11.25	164	281888	3.725	ppbv	92
53) Toluene	9.83	91	36446m	0.159	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL063020\
Data File : VL035140.D
Acq On : 1 Jul 2020 00:55
Operator : SY/AP
Sample : OC062920 CAN 10661
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
QC062920 CAN 10661

Manual Integrations
APPROVED

MMDadoda
7/2/2020 5:01:00 PM

Quant Time: Jul 01 05:09:56 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL061920AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Sat Jun
QLast Update : Sat Jun 20 02:10:36 2020
Response via : Initial Calibration

