

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL070518\
 Data File : VL032215.D
 Acq On : 5 Jul 2018 22:45
 Operator : SY/AP
 Sample : QC062918 CAN 10258
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 QC062918 CAN 10258

Manual Integrations
 APPROVED

MMDadoda
 7/9/2018 10:47:38 AM

Quant Time: Jul 06 11:34:11 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL062918AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Jun 29 2018
 QLast Update : Fri Jun 29 22:27:58 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.84	49	1384879	10.00	ppbv	-0.02
33) 1,4-Difluorobenzene	7.38	114	3666678	10.00	ppbv	-0.02
55) Chlorobenzene-d5	12.33	117	3746528	10.00	ppbv	-0.03

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.67	95	2896024	10.11	ppbv	-0.03
Spiked Amount	10.000	Range	65 - 135	Recovery	=	101.10%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL070518\
Data File : VL032215.D
Acq On : 5 Jul 2018 22:45
Operator : SY/AP
Sample : OC062918 CAN 10258
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
QC062918 CAN 10258

Manual Integrations
APPROVED

MMDadoda
7/9/2018 10:47:38 AM

Quant Time: Jul 06 11:34:11 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL062918AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Jun
QLast Update : Fri Jun 29 22:27:58 2018
Response via : Initial Calibration

