

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL070518\
 Data File : VL032222.D
 Acq On : 6 Jul 2018 3:08
 Operator : SY/AP
 Sample : J3762-15 0.03 PPB LOD
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 MDL-AIR-03-QT3-2018

Manual Integrations
 APPROVED

MMDadoda
 7/9/2018 10:47:54 AM

Quant Time: Jul 06 11:38:13 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL062918AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Jun 29 2018
 QLast Update : Fri Jun 29 22:27:58 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.84	49	1635462	10.00	ppbv	-0.02
33) 1,4-Difluorobenzene	7.38	114	4005172	10.00	ppbv	-0.02
55) Chlorobenzene-d5	12.34	117	3907399	10.00	ppbv	-0.02

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.67	95	2989796	10.01	ppbv	-0.03
Spiked Amount	10.000	Range	65 - 135	Recovery	=	100.10%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
5) Vinyl Chloride	3.34	62	3593m	0.03	ppbv	
32) 1,1,1-Trichloroethane	6.71	97	9599m	0.03	ppbv	
35) Carbon Tetrachloride	7.22	117	9413m	0.03	ppbv	
38) Trichloroethene	8.05	130	5430m	0.04	ppbv	
52) Tetrachloroethene	11.46	164	5395m	0.04	ppbv	
63) 1,1,2,2-Tetrachloroethane	13.94	83	14729m	0.04	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL070518\
Data File : VL032222.D
Acq On : 6 Jul 2018 3:08
Operator : SY/AP
Sample : J3762-15 0.03 PPB LOD
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
MDL-AIR-03-QT3-2018

Manual Integrations
APPROVED

MMDadoda
7/9/2018 10:47:54 AM

Quant Time: Jul 06 11:38:13 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL062918AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Jun
QLast Update : Fri Jun 29 22:27:58 2018
Response via : Initial Calibration

