

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL070820\
 Data File : VL035182.D
 Acq On : 8 Jul 2020 13:16
 Operator : SY/AP
 Sample : L3190-01DL2 1200X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 PERM-01DL

Quant Time: Jul 09 11:15:31 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL070820AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LWed Jul 08 07:05:23 2020
 QLast Update : Wed Jul 08 07:05:23 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.66	49	1210716	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.18	114	2832813	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.11	117	2601310	10.00	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	14.45	95	2130026	10.08	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	100.80%
Target Compounds						
						Ovalue
27) Carbon Disulfide	4.53	76	10698183	68.310	ppbv	# 79
36) Benzene	6.89	78	20441	0.100	ppbv	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL070820\
Data File : VL035182.D
Acq On : 8 Jul 2020 13:16
Operator : SY/AP
Sample : L3190-01DL2 1200X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
PERM-01DL

Quant Time: Jul 09 11:15:31 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL070820AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Jul 08 07:05:23 2020
QLast Update : Wed Jul 08 07:05:23 2020
Response via : Initial Calibration

