

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL071222\
 Data File : VL039459.D
 Acq On : 12 Jul 2022 22:24
 Operator : SY/AP
 Sample : QC062 CAN 10154
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 QC062 CAN 10154

Quant Time: Jul 13 01:41:51 2022
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL062922AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LWed Jul 06 02:15:38 2022
 QLast Update : Wed Jul 06 02:15:38 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.684	49	1485820	10.000	ppbv	0.03
33) 1,4-Difluorobenzene	7.192	114	4049561	10.000	ppbv	0.04
55) Chlorobenzene-d5	12.102	117	3656597	10.000	ppbv	0.04

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.429	95	2735057	10.083	ppbv	0.04
Spiked Amount	10.000	Range	65 - 135	Recovery	=	100.800%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL071222\
Data File : VL039459.D
Acq On : 12 Jul 2022 22:24
Operator : SY/AP
Sample : QC062 CAN 10154
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
QC062 CAN 10154

Quant Time: Jul 13 01:41:51 2022
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL062922AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LWed Jul 06 02:15:38 2022
QLast Update : Wed Jul 06 02:15:38 2022
Response via : Initial Calibration

