

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL071620\
 Data File : VL035194.D
 Acq On : 16 Jul 2020 15:13
 Operator : SY/AP
 Sample : L3318-02 600X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 CARBON-03

Quant Time: Jul 17 10:00:29 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL070820AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LWed Jul 08 07:05:23 2020
 QLast Update : Wed Jul 08 07:05:23 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.63	49	1189930	10.00	ppbv	-0.03
33) 1,4-Difluorobenzene	7.15	114	2789178	10.00	ppbv	-0.03
55) Chlorobenzene-d5	12.09	117	2519069	10.00	ppbv	-0.03

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.42	95	2093147	10.23	ppbv	-0.03
Spiked Amount	10.000	Range	65 - 135	Recovery	=	102.30%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
27) Carbon Disulfide	4.50	76	12571005	81.670	ppbv	# 67

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL071620\
Data File : VL035194.D
Acq On : 16 Jul 2020 15:13
Operator : SY/AP
Sample : L3318-02 600X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
CARBON-03

Quant Time: Jul 17 10:00:29 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL070820AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Jul 08 07:05:23 2020
QLast Update : Wed Jul 08 07:05:23 2020
Response via : Initial Calibration

