

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL071620\
 Data File : VL035204.D
 Acq On : 16 Jul 2020 21:42
 Operator : SY/AP
 Sample : L3318-01DUP 1200X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 PERM-02DUP

Quant Time: Jul 17 06:49:38 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL070820AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Jul 08 07:05:23 2020
 QLast Update : Wed Jul 08 07:05:23 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.63	49	1029376	10.00	ppbv	-0.02
33) 1,4-Difluorobenzene	7.15	114	2510065	10.00	ppbv	-0.03
55) Chlorobenzene-d5	12.09	117	2347296	10.00	ppbv	-0.03

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.42	95	1961542	10.29	ppbv	-0.03
Spiked Amount	10.000	Range	65 - 135	Recovery	=	102.90%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
5) Vinyl Chloride	3.20	62	3953m	0.058	ppbv	
13) Ethanol	3.57	45	3793m	0.277	ppbv	
15) Acetone	3.82	43	39006m	0.280	ppbv	
20) Methylene Chloride	4.30	84	20711	0.342	ppbv	# 89
27) Carbon Disulfide	4.51	76	6114698	45.922	ppbv	94
32) 1,1,1-Trichloroethane	6.48	97	8804m	0.055	ppbv	
35) Carbon Tetrachloride	7.00	117	9270m	0.057	ppbv	
38) Trichloroethene	7.83	130	3539m	0.048	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL071620\
Data File : VL035204.D
Acq On : 16 Jul 2020 21:42
Operator : SY/AP
Sample : L3318-01DUP 1200X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
PERM-02DUP

Quant Time: Jul 17 06:49:38 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL070820AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Jul 08 07:05:23 2020
QLast Update : Wed Jul 08 07:05:23 2020
Response via : Initial Calibration

