

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL072022\
 Data File : VL039513.D
 Acq On : 20 Jul 2022 16:38
 Operator : SY/AP
 Sample : N3750-02 40X
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 SV-2DL2

Quant Time: Jul 21 06:07:44 2022
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL062922AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LWed Jul 06 02:15:38 2022
 QLast Update : Wed Jul 06 02:15:38 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.684	49	905988	10.000	ppbv	0.03
33) 1,4-Difluorobenzene	7.192	114	2676672	10.000	ppbv	0.04
55) Chlorobenzene-d5	12.105	117	2297582	10.000	ppbv	0.04
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	14.433	95	1769201	10.380	ppbv	0.04
Spiked Amount	10.000	Range	65 - 135	Recovery	=	103.800%
Target Compounds						
29) Chloroform	5.767	83	2318143	15.048	ppbv	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL072022\
Data File : VL039513.D
Acq On : 20 Jul 2022 16:38
Operator : SY/AP
Sample : N3750-02 40X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
SV-2DL2

Quant Time: Jul 21 06:07:44 2022
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL062922AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LWed Jul 06 02:15:38 2022
QLast Update : Wed Jul 06 02:15:38 2022
Response via : Initial Calibration

