

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL072224\
 Data File : VL041384.D
 Acq On : 22 Jul 2024 13:22
 Operator : SY/MD
 Sample : VIBLK
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VIBLK

Quant Time: Jul 23 03:14:37 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL062624AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Tue Jul 23 02:50:01 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.189	49	971335	10.000	ppbv	-0.02
33) 1,4-Difluorobenzene	6.648	114	2637895	10.000	ppbv	-0.02
55) Chlorobenzene-d5	11.462	117	2306789	10.000	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	13.770	95	1753745	10.143	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	101.400%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL072224\
Data File : VL041384.D
Acq On : 22 Jul 2024 13:22
Operator : SY/MD
Sample : VIBLK
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VIBLK

Quant Time: Jul 23 03:14:37 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL062624AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Tue Jul 23 02:50:01 2024
Response via : Initial Calibration

