

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL072225\
 Data File : VL042732.D
 Acq On : 22 Jul 2025 15:41
 Operator : SY/MD
 Sample : VL0722ABL01
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VL0722ABL01

Quant Time: Jul 23 03:02:11 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL072225AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Wed Jul 23 02:51:39 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.790	49	99924	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.962	114	300308	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.888	117	270398	10.000	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	10.380	95	221143	10.119	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	101.200%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL072225\
Data File : VL042732.D
Acq On : 22 Jul 2025 15:41
Operator : SY/MD
Sample : VL0722ABL01
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VL0722ABL01

Quant Time: Jul 23 03:02:11 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL072225AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Wed Jul 23 02:51:39 2025
Response via : Initial Calibration

