

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL072225\
 Data File : VL042744.D
 Acq On : 22 Jul 2025 23:50
 Operator : SY/MD
 Sample : Q2632-02DL 10X
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 OA1DL

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 07/23/2025
 Supervised By :Mahesh Dadoda 07/23/2025

Quant Time: Jul 23 03:08:14 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL072225AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 23 2025
 QLast Update : Wed Jul 23 02:51:39 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.787	49	107003	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.959	114	308379	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.882	117	271224	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	10.377	95	222133	10.133	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	101.300%
Target Compounds						
					Qvalue	
9) Propene	1.486	41	1065	0.159	ppbv #	48
13) Ethanol	1.722	45	4899m	4.345	ppbv	
15) Acetone	1.839	43	11964	0.668	ppbv	94
20) Methylene Chloride	2.062	84	4345	0.771	ppbv	95
52) Tetrachloroethene	8.228	164	1445m	0.125	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL072225\
Data File : VL042744.D
Acq On : 22 Jul 2025 23:50
Operator : SY/MD
Sample : Q2632-02DL 10X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
OA1DL

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 07/23/2025
Supervised By :Mahesh Dadoda 07/23/2025

Quant Time: Jul 23 03:08:14 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL072225AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Wed Jul 23 02:51:39 2025
Response via : Initial Calibration

