

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL072425\
 Data File : VL042768.D
 Acq On : 24 Jul 2025 16:53
 Operator : SY/MD
 Sample : Q2671-02DL 10X
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 118-IA-1DL

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 07/25/2025
 Supervised By :Mahesh Dadoda 07/25/2025

Quant Time: Jul 25 01:29:12 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL072325AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2
 QLast Update : Thu Jul 24 05:24:41 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.794	49	151863	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.965	114	412621	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.888	117	353060	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	10.384	95	275055	9.837	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	98.400%
Target Compounds						
					Qvalue	
9) Propene	1.489	41	41514m	4.367	ppbv	
13) Ethanol	1.726	45	96968	138.940	ppbv	90
15) Acetone	1.839	43	50277	2.398	ppbv	97
19) Isopropyl Alcohol	1.891	45	27840	2.498	ppbv	93
26) Hexane	2.833	57	13749	0.675	ppbv #	52

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL072425\
Data File : VL042768.D
Acq On : 24 Jul 2025 16:53
Operator : SY/MD
Sample : Q2671-02DL 10X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
118-IA-1DL

Quant Time: Jul 25 01:29:12 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL072325AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Thu Jul 24 05:24:41 2025
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : Semsettin Yesilyurt 07/25/2025
Supervised By : Mahesh Dadoda 07/25/2025

