

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL072522\
 Data File : VL039527.D
 Acq On : 25 Jul 2022 22:07
 Operator : SY/AP
 Sample : N3842-02DL 40X
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 SV1DL

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 07/26/2022
 Supervised By :Mahesh Dadoda 07/26/2022

Quant Time: Jul 26 08:08:20 2022

Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL062922AIR.M

Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_LWed Jul 6

QLast Update : Wed Jul 06 02:15:38 2022

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.671	49	1132418	10.000	ppbv	0.02
33) 1,4-Difluorobenzene	7.179	114	3124828	10.000	ppbv	0.02
55) Chlorobenzene-d5	12.089	117	2718203	10.000	ppbv	0.02
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	14.417	95	2057185	10.202	ppbv	0.03
Spiked Amount	10.000	Range	65 - 135	Recovery	=	102.000%
Target Compounds						
					Qvalue	
9) Propene	3.028	41	15744	0.243	ppbv	87
13) Ethanol	3.623	45	3976m	1.056	ppbv	
15) Acetone	3.864	43	124820m	0.932	ppbv	
20) Methylene Chloride	4.349	84	71459m	1.338	ppbv	
23) Vinyl Acetate	5.076	43	24526m	0.274	ppbv	
26) Hexane	5.693	57	88225	0.797	ppbv #	37
52) Tetrachloroethene	11.224	164	30078m	0.309	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL072522\
Data File : VL039527.D
Acq On : 25 Jul 2022 22:07
Operator : SY/AP
Sample : N3842-02DL 40X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
SV1DL

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 07/26/2022
Supervised By :Mahesh Dadoda 07/26/2022

Quant Time: Jul 26 08:08:20 2022
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL062922AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LWed Jul
QLast Update : Wed Jul 06 02:15:38 2022
Response via : Initial Calibration

