

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL072524\
 Data File : VL041399.D
 Acq On : 25 Jul 2024 13:59
 Operator : SY/MD
 Sample : VSTDICCC010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICCC010

Quant Time: Jul 26 05:05:14 2024

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL072524AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Fri Jul 26 04:51:44 2024

Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : Semsettin Yesilyurt 07/26/2024
 Supervised By : Mahesh Dadoda 07/29/2024

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.176	49	1227171	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	6.645	114	3353351	10.000	ppbv	0.00
55) Chlorobenzene-d5	11.465	117	3260833m	10.000	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene 13.770 95 2550034 10.170 ppbv 0.00
 Spiked Amount 10.000 Range 65 - 135 Recovery = 101.700%

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	2.706	85	1748662	8.150	ppbv	100
3) Chlorodifluoromethane	2.652	51	1312868	9.025	ppbv	100
4) Chloromethane	2.790	50	692256	9.024	ppbv	99
5) Vinyl Chloride	2.899	62	699389	9.214	ppbv	100
6) Bromomethane	3.099	94	361665	9.227	ppbv	100
7) Chloroethane	3.176	64	248340	9.630	ppbv	100
8) Dichlorotetrafluoroethane	2.835	85	1694634	9.039	ppbv	100
9) Propene	2.671	41	600396	8.889	ppbv	100
10) Heptane	7.568	43	1472225	9.801	ppbv	99
11) Trichlorofluoromethane	3.542	101	1696907	9.145	ppbv	100
12) 1,1,2-Trichlorotrifluo...	4.047	101	1226328	9.136	ppbv	100
13) Ethanol	3.224	45	43077m	7.073	ppbv	
14) Bromoethene	3.346	108	505960	9.786	ppbv	100
15) Acetone	3.446	43	1222864	8.859	ppbv	100
16) 1,3-Butadiene	2.964	39	681843	9.269	ppbv	100
17) tert-Butyl alcohol	3.861	59	1291681	10.276	ppbv	100
18) 1,1-Dichloroethene	3.861	96	558470	9.575	ppbv	100
19) Isopropyl Alcohol	3.555	45	792740	10.155	ppbv	100
20) Methylene Chloride	3.915	84	455611	8.929	ppbv	100
21) Allyl Chloride	3.980	41	890718	9.827	ppbv	100
22) trans-1,2-Dichloroethene	4.430	96	549902	9.711	ppbv	100
23) Vinyl Acetate	4.610	43	1085315	10.104	ppbv	100
24) 1,1-Dichloroethane	4.549	63	1154644	9.415	ppbv	100
25) Ethyl Acetate	5.189	43	2597295	9.430	ppbv	100
26) Hexane	5.198	57	1090849	9.420	ppbv	100
27) Carbon Disulfide	4.105	76	1375021	10.384	ppbv	99
28) Methyl tert-Butyl Ether	4.568	73	681033	9.329	ppbv	100
29) Chloroform	5.259	83	1735318	9.163	ppbv	100
30) Cyclohexane	6.590	84	937363	9.878	ppbv	100
31) cis-1,2-Dichloroethene	5.063	61	1091838	9.548	ppbv	100
32) 1,1,1-Trichloroethane	5.989	97	1715896	9.483	ppbv	100
34) 2-Butanone	4.771	43	1567685	9.654	ppbv	100
35) Carbon Tetrachloride	6.484	117	1836059	9.983	ppbv	100
36) Benzene	6.359	78	2363293	9.591	ppbv	100
37) 1,2-Dichloroethane	5.796	62	1317693	9.664	ppbv	99
38) Trichloroethene	7.282	130	1074518	9.605	ppbv	100
39) 1,2-Dichloropropane	7.063	63	891517	9.403	ppbv	100
40) 1,4-Dioxane	7.275	88	384473	11.177	ppbv	100
41) Tetrahydrofuran	5.542	42	931064	9.892	ppbv	100
42) Bromodichloromethane	7.240	83	1932798	9.793	ppbv	100
43) Methyl Methacrylate	7.465	69	948521	10.646	ppbv	100
44) 2,2,4-Trimethylpentane	7.317	57	3907968	9.554	ppbv	100
45) t-1,3-Dichloropropene	8.703	75	963383	11.835	ppbv	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL072524\
 Data File : VL041399.D
 Acq On : 25 Jul 2024 13:59
 Operator : SY/MD
 Sample : VSTDICCC010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICCC010

Manual Integrations
 APPROVED

Reviewed By : Semsettin Yesilyurt 07/26/2024
 Supervised By : Mahesh Dadoda 07/29/2024

Quant Time: Jul 26 05:05:14 2024

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL072524AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Fri Jul 26 04:51:44 2024

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) cis-1,3-Dichloropropene	8.134	75	1281640	10.855	ppbv	100
47) 1,1,2-Trichloroethane	8.893	97	987016	9.602	ppbv	100
48) Dibromochloromethane	9.706	129	1692234	10.224	ppbv	100
49) Bromoform	12.404	173	1496236	10.670	ppbv	99
50) 4-Methyl-2-Pentanone	8.172	43	2603717	10.051	ppbv	100
51) 2-Hexanone	9.542	43	2141345	10.819	ppbv	100
52) Tetrachloroethene	10.613	164	907950	9.585	ppbv	100
53) Toluene	9.214	91	2779504	10.036	ppbv	100
54) 1,2-Dibromoethane	10.002	107	1438535	10.342	ppbv	100
56) 1,1,1,2-Tetrachloroethane	11.503	131	1213659	9.640	ppbv #	58
57) Chlorobenzene	11.523	112	2174265	9.343	ppbv	100
58) Ethyl Benzene	12.085	91	3752771	9.870	ppbv	100
59) m/p-Xylene	12.368	91	6382118m	19.609	ppbv	
60) o-Xylene	13.053	91	3036791	9.732	ppbv	100
61) Styrene	12.896	104	1513410	10.748	ppbv	100
62) Isopropylbenzene	14.024	105	4462038	9.595	ppbv	100
63) 1,1,2,2-Tetrachloroethane	13.040	83	1853886	9.567	ppbv	100
64) n-propylbenzene	14.912	120	1205437	9.895	ppbv	100
65) tert-Butylbenzene	16.085	119	4027165	9.564	ppbv	100
66) Benzyl Chloride	16.323	91	299511	11.162	ppbv	100
67) sec-Butylbenzene	16.635	105	5605419	9.602	ppbv	100
69) p-Isopropyltoluene	16.989	119	4737222	9.713	ppbv	100
70) n-Butylbenzene	17.883	91	4851380	9.796	ppbv	100
71) 2-Chlorotoluene	14.799	91	3368997	9.707	ppbv	100
72) 4-Ethyltoluene	15.188	105	3711228	10.293	ppbv	99
73) 1,3,5-Trimethylbenzene	15.346	105	3175291	10.049	ppbv	100
74) 1,2,4-Trimethylbenzene	16.105	105	3502689	9.621	ppbv	100
75) 1,3-Dichlorobenzene	16.330	146	2192883	9.642	ppbv	100
76) 1,4-Dichlorobenzene	16.471	146	2140459	9.596	ppbv	100
77) 1,2-Dichlorobenzene	17.140	146	2131506	9.501	ppbv	100
78) Hexachloro-1,3-Butadiene	22.628	225	1626038	8.347	ppbv	100
79) Naphthalene	21.381	128	3669460	11.481	ppbv	100
80) Naphthalene,2-methyl-	24.371	142	3235986	12.454	ppbv	100
81) 1,2,4-Trichlorobenzene	21.133	180	1912951	9.962	ppbv	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

