

Data Path : W:\HPCHEM1\MSVOA_L\DATA\VL072715\
 Data File : VL025758.D
 Acq On : 27 Jul 2015 18:13
 Operator : SY/MD
 Sample : G3092-01DL 40X
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 A-221-BDL

Manual Integrations
 APPROVED

sam
 7/28/2015 12:00:24 PM

Quant Time: Jul 28 02:03:34 2015
 Quant Method : W:\HPCHEM1\MSVOA_L\METHODS\VL070715AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Jul 08 02:03:34 2015
 QLast Update : Wed Jul 08 02:22:55 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	6.65	49	1018765	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	8.32	114	2692425	10.00	ppbv	0.00
55) Chlorobenzene-d5	13.75	117	3241149	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	16.29	95	2736358	10.01	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	100.10%

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
9) Propene	3.60	41	9105m	0.19	ppbv	
13) Ethanol	4.36	45	11431	1.54	ppbv	96
15) Acetone	4.64	43	46676	0.26	ppbv #	64
20) Methylene Chloride	5.14	84	14645	0.20	ppbv #	89
52) Tetrachloroethene	12.80	164	3794563	29.34	ppbv	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA_L\DATA\VL072715\
Data File : VL025758.D
Acq On : 27 Jul 2015 18:13
Operator : SY/MD
Sample : G3092-01DL 40X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
A-221-BDL

Manual Integrations
APPROVED

sam
7/28/2015 12:00:24 PM

Quant Time: Jul 28 02:03:34 2015
Quant Method : W:\HPCHEM1\MSVOA_L\METHODS\VL070715AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Jul
QLast Update : Wed Jul 08 02:22:55 2015
Response via : Initial Calibration

