

Data Path : W:\HPCHEM1\MSVOA_L\DATA\VL072715\
 Data File : VL025762.D
 Acq On : 27 Jul 2015 20:57
 Operator : SY/MD
 Sample : G3092-03DL 40X
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 A-217-BDL

Manual Integrations
 APPROVED

MMDadoda
 7/28/2015 12:57:19 PM

Quant Time: Jul 28 11:27:38 2015
 Quant Method : W:\HPCHEM1\MSVOA_L\METHODS\VL070715AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Jul 08 2015
 QLast Update : Wed Jul 08 02:22:55 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	6.65	49	731097	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	8.32	114	2003082	10.00	ppbv	0.00
55) Chlorobenzene-d5	13.74	117	2475896	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	16.28	95	2092798	10.02	ppbv	-0.01
Spiked Amount	10.000	Range	65 - 135	Recovery	=	100.20%

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
9) Propene	3.60	41	16176m	0.47	ppbv	
13) Ethanol	4.36	45	9893m	1.86	ppbv	
15) Acetone	4.64	43	62620	0.48	ppbv #	83
52) Tetrachloroethene	12.78	164	8174	0.08	ppbv #	78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA_L\DATA\VL072715\
Data File : VL025762.D
Acq On : 27 Jul 2015 20:57
Operator : SY/MD
Sample : G3092-03DL 40X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
A-217-BDL

Manual Integrations
APPROVED

MMDadoda
7/28/2015 12:57:19 PM

Quant Time: Jul 28 11:27:38 2015
Quant Method : W:\HPCHEM1\MSVOA_L\METHODS\VL070715AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Jul
QLast Update : Wed Jul 08 02:22:55 2015
Response via : Initial Calibration

