

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL072920\
 Data File : VL035244.D
 Acq On : 29 Jul 2020 13:36
 Operator : SY/AP
 Sample : L3474-01DL 10X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 1-VOA

Manual Integrations
 APPROVED

MMDadoda
 7/30/2020 6:10:45 PM

Quant Time: Jul 30 11:09:39 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL072920AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Jul 29 2020
 QLast Update : Wed Jul 29 07:14:29 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.65	49	1398145	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.18	114	3283865	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.11	117	3021879	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.45	95	2413553	10.00	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	100.00%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
9) Propene	2.97	41	23978m	0.444	ppbv	
13) Ethanol	3.58	45	1114939	50.778	ppbv	96
15) Acetone	3.81	43	3950282	21.182	ppbv	97
19) Isopropyl Alcohol	3.95	45	10947791	89.008	ppbv #	97
20) Methylene Chloride	4.32	84	604509	7.121	ppbv	99
26) Hexane	5.68	57	901295	6.860	ppbv #	42
28) Methyl tert-Butyl Ether	5.01	73	46715	0.249	ppbv	100
29) Chloroform	5.74	83	668020	3.032	ppbv	99
53) Toluene	9.83	91	65515	0.270	ppbv	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL072920\
Data File : VL035244.D
Acq On : 29 Jul 2020 13:36
Operator : SY/AP
Sample : L3474-01DL 10X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
1-VOA

Quant Time: Jul 30 11:09:39 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL072920AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Jul
QLast Update : Wed Jul 29 07:14:29 2020
Response via : Initial Calibration

Manual Integrations
APPROVED

MMDadoda
7/30/2020 6:10:45 PM

