

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL072920\
 Data File : VL035245.D
 Acq On : 29 Jul 2020 14:14
 Operator : SY/AP
 Sample : L3474-01DL DUP 10X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 1-VOADUP

Manual Integrations
 APPROVED

MMDadoda
 7/30/2020 6:10:47 PM

Quant Time: Jul 30 11:10:31 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL072920AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Jul 29 2020
 QLast Update : Wed Jul 29 07:14:29 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.66	49	1353509	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.18	114	3161548	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.11	117	2976138	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.45	95	2377649	10.00	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	100.00%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
9) Propene	2.98	41	25183m	0.482	ppbv	
13) Ethanol	3.58	45	1181667	55.592	ppbv	94
15) Acetone	3.81	43	4218933	23.368	ppbv	97
19) Isopropyl Alcohol	3.95	45	11536643	96.889	ppbv #	97
20) Methylene Chloride	4.32	84	666187	8.106	ppbv	95
26) Hexane	5.68	57	970005	7.626	ppbv #	42
28) Methyl tert-Butyl Ether	5.01	73	42927m	0.236	ppbv	
29) Chloroform	5.74	83	701457	3.288	ppbv	98
53) Toluene	9.83	91	70096m	0.300	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL072920\
Data File : VL035245.D
Acq On : 29 Jul 2020 14:14
Operator : SY/AP
Sample : L3474-01DLUP 10X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
1-VOADUP

Quant Time: Jul 30 11:10:31 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL072920AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Jul
QLast Update : Wed Jul 29 07:14:29 2020
Response via : Initial Calibration

Manual Integrations
APPROVED

MMDadoda
7/30/2020 6:10:47 PM

