

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL072920\
 Data File : VL035251.D
 Acq On : 29 Jul 2020 18:09
 Operator : SY/AP
 Sample : L3474-03DL 10X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 SINKDL

Manual Integrations
 APPROVED

MMDadoda
 7/30/2020 6:11:18 PM

Quant Time: Jul 30 04:35:50 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL072920AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Wed Jul 29 2020
 QLast Update : Wed Jul 29 07:14:29 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.66	49	1386803	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.19	114	3183201	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.12	117	2904910	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.46	95	2381179	10.26	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	102.60%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
9) Propene	2.98	41	18301m	0.342	ppbv	
13) Ethanol	3.58	45	2331578	107.056	ppbv	95
15) Acetone	3.82	43	12358546	66.810	ppbv #	68
19) Isopropyl Alcohol	3.96	45	15622661	128.055	ppbv #	97
20) Methylene Chloride	4.33	84	415951	4.940	ppbv	99
26) Hexane	5.68	57	2203108	16.906	ppbv #	42
28) Methyl tert-Butyl Ether	5.02	73	352745	1.893	ppbv	99
29) Chloroform	5.75	83	747241	3.419	ppbv	99
53) Toluene	9.83	91	56885	0.242	ppbv	94
59) m/p-Xylene	13.01	91	68461m	0.239	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL072920\
Data File : VL035251.D
Acq On : 29 Jul 2020 18:09
Operator : SY/AP
Sample : L3474-03DL 10X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
SINKDL

Manual Integrations
APPROVED

MMDadoda
7/30/2020 6:11:18 PM

Quant Time: Jul 30 04:35:50 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL072920AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Jul
QLast Update : Wed Jul 29 07:14:29 2020
Response via : Initial Calibration

