

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL080118\
 Data File : VL032333.D
 Acq On : 1 Aug 2018 21:55
 Operator : SY/AP
 Sample : J4235-01DL 10X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 MARKETING-CONFERENCE-ROOM

Manual Integrations
 APPROVED

MMDadoda
 8/2/2018 2:04:58 PM

Quant Time: Aug 02 02:05:16 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL072718AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Jul 27 2018
 QLast Update : Fri Jul 27 07:05:37 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.83	49	1945629	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.36	114	4391190	10.00	ppbv	-0.01
55) Chlorobenzene-d5	12.32	117	4671388	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.66	95	3680450	10.15	ppbv	-0.01
Spiked Amount	10.000	Range	65 - 135	Recovery	=	101.50%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
9) Propene	3.09	41	28727m	0.33	ppbv	
13) Ethanol	3.70	45	329309	13.10	ppbv	100
15) Acetone	3.96	43	307052	1.67	ppbv	95
20) Methylene Chloride	4.47	84	37717m	0.51	ppbv	
26) Hexane	5.85	57	66301m	0.36	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL080118\
Data File : VL032333.D
Acq On : 1 Aug 2018 21:55
Operator : SY/AP
Sample : J4235-01DL 10X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
MARKETING-CONFERENCE-ROOM

Manual Integrations
APPROVED

MMDadoda
8/2/2018 2:04:58 PM

Quant Time: Aug 02 02:05:16 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL072718AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Jul
QLast Update : Fri Jul 27 07:05:37 2018
Response via : Initial Calibration

