

Data Path : Z:\voasrv\HPCHEM1\MSVOA L\Data\VL080118\
 Data File : VL032340.D
 Acq On : 2 Aug 2018 2:39
 Operator : SY/AP
 Sample : J4258-02DL2 40X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 SV-1DL2

Manual Integrations
 APPROVED

apatel
 8/14/2018 2:01:18 PM

Quant Time: Aug 02 02:07:12 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL072718AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Jul 27 2018
 QLast Update : Fri Jul 27 07:05:37 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.83	49	1813500	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.36	114	4525603	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.32	117	4423969	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.66	95	3519060	10.25	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	102.50%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
9) Propene	3.08	41	808344	9.915	ppbv	93
13) Ethanol	3.70	45	109564	4.676	ppbv	100
15) Acetone	3.95	43	7017765	41.058	ppbv	92
26) Hexane	5.84	57	49976	0.288	ppbv #	1
27) Carbon Disulfide	4.68	76	41181	0.261	ppbv #	81
34) 2-Butanone	5.38	43	4619152	15.769	ppbv	99
36) Benzene	7.08	78	113785m	0.285	ppbv	
51) 2-Hexanone	10.36	43	690589	1.753	ppbv #	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA L\Data\VL080118\
Data File : VL032340.D
Acq On : 2 Aug 2018 2:39
Operator : SY/AP
Sample : J4258-02DL2 40X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
SV-1DL2

Manual Integrations
APPROVED

apatel
8/14/2018 2:01:18 PM

Quant Time: Aug 02 02:07:12 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL072718AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Jul
QLast Update : Fri Jul 27 07:05:37 2018
Response via : Initial Calibration

