

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL080320\
 Data File : VL035319.D
 Acq On : 3 Aug 2020 14:51
 Operator : SY/AP
 Sample : VSTDIC0.1
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDIC0.1

Manual Integrations
 APPROVED

apatel
 8/6/2020 2:44:46 PM

Quant Time: Aug 03 15:23:41 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL080320AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Mon Aug 03 15:23:41 2020
 QLast Update : Mon Aug 03 15:02:12 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.65	49	1409508	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.17	114	3498308	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.11	117	3318035	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.44	95	2823526	10.65	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	106.50%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
5) Vinyl Chloride	3.21	62	8913	0.097	ppbv	# 88
32) 1,1,1-Trichloroethane	6.50	97	22806m	0.100	ppbv	
35) Carbon Tetrachloride	7.02	117	19878m	0.085	ppbv	
38) Trichloroethene	7.83	130	8949m	0.089	ppbv	
52) Tetrachloroethene	11.25	164	7507m	0.081	ppbv	
63) 1,1,2,2-Tetrachloroethane	13.71	83	22396m	0.080	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL080320\
Data File : VL035319.D
Acq On : 3 Aug 2020 14:51
Operator : SY/AP
Sample : VSTDIC0.1
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VSTDIC0.1

Manual Integrations
APPROVED

apatel
8/6/2020 2:44:46 PM

Quant Time: Aug 03 15:23:41 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL080320AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Mon Aug
QLast Update : Mon Aug 03 15:02:12 2020
Response via : Initial Calibration

