

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_L\DATA\VL080320\
 Data File : VL035320.D
 Acq On : 3 Aug 2020 15:28
 Operator : SY/AP
 Sample : VSTDICC0.03
 Misc : 400ml/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC0.03

Manual Integrations
 APPROVED

apatel
 8/6/2020 2:44:53 PM

Quant Time: Aug 03 16:26:09 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL080320AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Mon Aug 03 16:26:09 2020
 QLast Update : Mon Aug 03 15:02:12 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.65	49	1351846	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.17	114	3350817	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.11	117	3205264	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.45	95	2726701	10.65	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	106.50%

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
5) Vinyl Chloride	3.20	62	2648m	0.030	ppbv	
32) 1,1,1-Trichloroethane	6.50	97	6566m	0.030	ppbv	
35) Carbon Tetrachloride	7.02	117	6615m	0.029	ppbv	
38) Trichloroethene	7.83	130	1411m	0.015	ppbv	
52) Tetrachloroethene	11.25	164	2202m	0.025	ppbv	
63) 1,1,2,2-Tetrachloroethane	13.71	83	7293m	0.027	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_L\DATA\VL080320\
Data File : VL035320.D
Acq On : 3 Aug 2020 15:28
Operator : SY/AP
Sample : VSTDIC0.03
Misc : 400ml/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VSTDIC0.03

Manual Integrations
APPROVED

apatel
8/6/2020 2:44:53 PM

Quant Time: Aug 03 16:26:09 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_L\METHODS\VL080320AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Mon Aug
QLast Update : Mon Aug 03 15:02:12 2020
Response via : Initial Calibration

